CORRIGENDUM

Sl No.		To be read as:
	Technical specifications	
7	The Gantry should have 3D Positioning Laser lights, with a patient observation camera.	The Gantry should have 3D Positioning Laser lights. A patient observation camera should be provided integrated into the gantry or within the scanner room with viewing facility in the console room.
23	Field of view should be 50cm.	43cm or better. The maximum available FOV for the quoted model should be offered.
24	Detectors:24 High contrast spatial resolution at 0%MTF with full FOV >15lp/ cm or higher	Detectors:24 High contrast spatial resolution at 0%MTF with full FOV >15lp/ cm or higher. Better configuration for the model quoted should be offered.
32	The patient table offered should have a minimum load bearing capacity of at least 200 kg.	The patient table offered should have a minimum load bearing capacity of at least 150kg or better.
33	Patient Table: The Floating tabletop width should be at least 45 cm or better for better comfort vertical range of travel should be at least 40 cm or better patient comfort.	The tabletop width should be at least 40 cm or better for better comfort. Vertical range of travel should be at least 40 cm or better patient comfort. The maximum available table top width with the quoted model should be offered.
62	Uptime: 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.	Uptime: 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty to extend CMC period by double the downtime period
	Accessories	
63	63 OEM Post Processing Workstations for CT: a. Two standalone multi-modality independent workstations with same user interface as the console computer from the same manufacturer as the CT Scanner.	63 OEM Post Processing Workstations for CT: a. Two standalone multi-modality independent workstations with same user interface as the console computer, both from the same manufacturer as the CT Scanner.
63	OEM Post Processing Workstations for CT: e. The workstations should be multivendor, multimodality, multiuser compatible.	OEM Post Processing Workstations for CT: e. The workstations should be multivendor, multimodality compatible.

Sl No.		To be read as:
66	Dual Head Contrast Injector of reputed make with compatible 200 syringes and 500 Y-tubings.	Dual Head Contrast Injector of reputed make with compatible 200 syringes and 500 Y-tubings. The model should be a latest model. Capability to load 500ml vials / contrast pouches is preferred.
	Turnkey	
80	Room shielding to be done by seller as per AERB guidelines.	Room shielding to be done by seller as per AERB guidelines. TLD badge management system using RFID based technology for 40 users is to be provided in the room designated by buyer.
	Buyer added bid specific ATC	
16.5	Model release date in India should not be older than five years from date of notification of tender.	Latest model / software version shall be quoted and supplied. The OEM should give an undertaking that 'All hardware and software updates and advanced applications for the model supplied, will be provided during the period of warranty and CMC.'
	Generic	
10	Training of users by an application engineer in machine operation, troubleshooting aspects and basic maintenance shall be provided once a year during the period of warranty.	Onsite Training of users by an application engineer in machine operation, troubleshooting aspects and basic maintenance to be provided for the period of at least 5 working days, every year during the warranty and CMC.
	MII Purchase preference	
	Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time).	All vendors may participate. However, Class 1/Class 2 local suppliers will be preferred as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020 (as amended and applicable time to time).
	In case of goods, 100% payment will be released within ten (10) days of issue of consignee receipt-cum-acceptance certificate (CRAC).	100% Payment shall be released through confirmed & irrevocable Inland Letter of Credit which is as under: 80% of Contract Value(Equipment & Local Accessoreis) payable on submission of documents: Delivery Challan/ Invoice/Insurance Certificate & Copy of Lorry Receipt and Balance 20% against issuance of satisfactory installation certificate by user.

Sl No.		To be read as:
	Additional clause for CMC	
3	GST shall be included in the CMC Charges quoted.	CMC charges including GST and applicable GST may be reimbursed.
	Consignee details	
	Delivery: 45 days	Delivery of the scanner should be done within 30 days from date of purchase order. Installation and full commissioning should be done within 90 days of receipt of stores/ goods at site.
	Vendors seeking exemption from EMD to submit all suitable evidence for evaluation.	

Seller Query/Representation

Id	Seller Query/Representation	Buyer Response
14	Detectors:24 High contrast spatial resolution at 0%MTF with full FOV >15lp/cm or higher modified to Detectors:24 High contrast spatial resolution at 2%MTF >15lp/cm or higher	24 High contrast spatial resolution at 0%MTF with full FOV >15lp/cm or higher. Better specifications in the model quoted should be offered.
15	Patient Table: 33. The Floating tabletop width should be at least 45 cm or better for better comfort vertical range of travel should be at least 40 cm or better patient comfort. Patient Table: 33. The tabletop width should be at least 40 cm or better for better comfort. vertical range of travel should be at least 40 cm or better patient comfort.	Modified to 'The tabletop width should be at least 40 cm or better for better comfort . vertical range of travel should be at least 40 cm or better patient comfort. The maximum available table top width with the model should be offered.'
16	63 OEM Post Processing Workstations for CT: e. The workstations should be multivendor, multimodality, multiuser compatible. 63 OEM Post Processing Workstations for CT: e. The workstations should be multivendor, multimodality compatible.	Modified to: 63 OEM Post Processing Workstations for CT: e. The workstations should be multivendor, multimodality compatible.
17	Tender Clause: Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time) Amendment Requested: Please delete the condition	Modified to 'class 1/Class 2 local suppliers preferred as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time)'
18	Model release date in India should not be older than five years from date of notification of tender	Modified to 'Latest model / software version shall be quoted and supplied. The OEM should give an undertaking that 'All hardware and software updates and advanced applications for the model supplied, will be provided during the period of warranty and CMC'.
19	Kindly confirm that SHPL is exempt from submission of Earnest Money Deposit (EMD) under the present tender, being a GeM Vendor Assessment (VA) verified OEM for the CT Scan category.	Tender Clause: EMD Detail Advisory Bank: State Bank of India EMD Amount:3000000 Any OEM seeking exemption must support the claim with valid documents

20	Tender Clause: Delivery 45 Days. Delivery & installation date will be 120 days from date of handing over of clear site or Issuance of Inland Letter of Credit or Site handover with permanent power availability or AERB site approval or statutory & regulatory approval , whichever is later.	Modified to: Delivery of the scanner should be done within 30 days from date of purchase order. Installation and full commissioning should be done within 90 days of receipt of stores/goods at site.
21	Training of users by an application engineer in machine operation, troubleshooting aspects and basic maintenance shall be provided once a year during the period of warranty. Training of users by an application engineer in machine operation, troubleshooting aspects and basic maintenance shall be provided once a year for 5 days and to 5 employees during the period of warranty.	Modified to: Onsite Training of users by an application engineer in machine operation, troubleshooting aspects and basic maintenance to be provided for the period of at least 5 working days, every year during the warranty and CMC
22	Atleast 95% Uptime during warranty and CMC Period	Modified to: 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty to extend CMC period by double the downtime period
23	In case of goods, 100% payment will be released within ten (10) days of issue of consignee receipt-cum-acceptance certificate (CRAC). 100% Payment shall be released through confirmed & irrevocable Inland Letter of Credit which is as under: 80% of Contract Value(Equipment & Local Accessoreis) payable on submission of documents: Delivery Challan/ Invoice/Insurance Certificate & Copy of Lorry Receipt and Balance 20% against issuance of satisfactory installation certificate by user.	Modified to: 100% Payment shall be released through confirmed & irrevocable Inland Letter of Credit which is as under: 80% of Contract Value(Equipment & Local Accessoreis) payable on submission of documents: Delivery Challan/ Invoice/Insurance Certificate & Copy of Lorry Receipt and Balance 20% against issuance of satisfactory installation certificate by user.
24	The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 50 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 50% of the contracted quantity during the currency of the contract at the contracted rates. "The purchaser reserves the right to increase/decrease the ordered quantity by up to [50] per cent at any time, till the final delivery date or within the validity of the Bid offer, whichever is earlier by giving reasonable notice even though the quantity ordered initially has been supplied in full before the last date of the	Not modified

	delivery period.	
25	Point (Pg,09 Gem tender): GST shall be included in the CMC Charges quoted.	CMC charges including GST and applicable GST may be reimbursed.
26	63 OEM Post Processing Workstations for CT: a. Two standalone multi-modality independent workstations with same user interface as the console computer from the same manufacturer as the CT Scanner	63 OEM Post Processing Workstations for CT: a. Two standalone multi-modality independent workstations from OEM, with same user interface as the console computer from the same manufacturer as the CT Scanner